

Instructions for use of the Homepump Eclipse™

Supplies:

- ✓ One (1) Homepump Eclipse™ of room temperature medication
 - Do not microwave or put in warm water
 - Check patient's name, medication and directions
- ✓ Four (4) alcohol wipes
- ✓ Two (2) 10mL NS flush (normal saline syringe)
- ✓ One (1) SwabCap (orange, green or blue cap)
- ✓ One (1) heparin flush (if instructed)

Steps:

1. **WASH HANDS**
2. **CLEAN** area where you will be working
3. **CLEAN** IV line with an alcohol wipe for 30 seconds
4. **REMOVE AIR** from NS flush and **FLUSH** IV line with 10mL NS
5. **CLEAN** IV line with an alcohol wipe for 30 seconds
6. **REMOVE** cap from medication tubing
7. **CONNECT** medication ball to IV line
8. **UNCLAMP** medication ball
9. **INFUSE** medication
10. **WASH HANDS** when Eclipse is **EMPTY**
11. **REMOVE** medication ball from IV line
12. **CLEAN** IV line with an alcohol wipe for 30 seconds
13. **REMOVE AIR** from NS flush and **FLUSH** IV line with 10mL
14. If instructed, **REMOVE AIR** from heparin flush and **FLUSH** IV line
15. **PLACE** SwabCap on end of IV line
16. **CLAMP** IV line (if instructed)
17. **THROW AWAY** all used supplies

**Training video
available!**




Scan the QR code using the camera on your smart phone or tablet to watch a quick instructional video.

SASH Teaching Guide



S

Saline




A

Administration
of Medicine




S

Saline



H

Heparin
(if ordered)



Remember to

- Use a new alcohol wipe for each step and scrub for at least 30 seconds.
- Clean your hands thoroughly using antibacterial soap and warm water. If no soap and water is available, use an alcohol-based hand sanitizer.
- If required, remove medication from refrigerator prior to use. Refer to medication label for required time.

